

**Kit CoA Cover Page**

KAPA Pure Beads (5ml)			
Kit Code	Part Number	Lot Number	Kit Expiry Date
KK8000	07983271001	004266-27-1	2017-12-02

Component Code	Component Description	Component Lot Number
KS8000	KAPA Pure Beads (5 ml)	00066318

*CoA's are not issued for complete kits, but for the individual component lots from which kits are assembled. CoA's for all component lots listed are attached.*

Generated By	Date
Ayesha Mohamed (QC Scientist)	2016-09-09



## Certificate of Analysis

### PRODUCT DETAILS

Product name	KAPA Pure Beads	
Code & Lot number	KS8000	66318
Pack size	5 mL	
Bulk Code & Lot number	BS0002	66204

### QUALITY CONTROL PARAMETERS

Parameter	Specification	Result
Functional assay	KAPA Pure Beads are functionally tested by capturing fragmented plasmid DNA at a 1.2X clean-up ratio. Fragment capture profile is determined by gel electrophoresis.	Passed
DNA contamination	A standard KAPA SYBR FAST no template reaction with KAPA Pure Beads (buffer only) contains <45 fg/μL bacterial genomic DNA ( <i>E. coli</i> and related strains); as assessed by amplification of a 217 bp 16S rRNA fragment using a multicopy primer set in a 40-cycle reaction, no detectable human genomic DNA (as assessed by amplification of a 290 bp b-actin fragment using a multicopy primer set in a 40-cycle reaction) and no detectable adapter-ligated library DNA (as assessed by amplification using primers specific for the Illumina® TruSeq™ adapters or Ion Torrent™ adapters).	Passed

Generated by Toni Marinus (QC Scientist)

2016-04-07