

Kit CoA Cover Page

| KAPA Pure Beads (30ml) | | | |
|------------------------|-------------|------------|-----------------|
| Kit Code | Part Number | Lot Number | Kit Expiry Date |
| KK8001 | 07983280001 | 003834-2-1 | 2017-08-28 |

| Component Code | Component Description | Component Lot Number |
|----------------|-------------------------|----------------------|
| KS8001 | KAPA Pure Beads (30 ml) | 00066317 |

CoA's are not issued for complete kits, but for the individual component lots from which kits are assembled. CoA's for all component lots listed are attached.

| Generated By | Date |
|--|------------|
| Chanell Herfurth (QC Stability Supervisor) | 2016-05-30 |



Certificate of Analysis

PRODUCT DETAILS

| | | |
|------------------------|-----------------|-------|
| Product name | KAPA Pure Beads | |
| Code & Lot number | KS8001 | 66317 |
| Pack size | 30 mL | |
| Bulk Code & Lot number | BS0002 | 66204 |

QUALITY CONTROL PARAMETERS

| Parameter | Specification | Result |
|-------------------|--|--------|
| Functional assay | KAPA Pure Beads are functionally tested by capturing fragmented plasmid DNA at a 1.2X clean-up ratio. Fragment capture profile is determined by gel electrophoresis. | Passed |
| DNA contamination | A standard KAPA SYBR FAST no template reaction with KAPA Pure Beads (buffer only) contains <45 fg/μL bacterial genomic DNA (<i>E. coli</i> and related strains); as assessed by amplification of a 217 bp 16S rRNA fragment using a multicopy primer set in a 40-cycle reaction, no detectable human genomic DNA (as assessed by amplification of a 290 bp b-actin fragment using a multicopy primer set in a 40-cycle reaction) and no detectable adapter-ligated library DNA (as assessed by amplification using primers specific for the Illumina® TruSeq™ adapters or Ion Torrent™ adapters). | Passed |

Generated by Toni Marinus (QC Scientist)

2016-04-07