

Kit CoA Cover Page

KAPA Pure Beads (5ml)			
Kit Code	Part Number	Lot Number	Kit Expiry Date
KK8000	07983271001	004480-1-1	2018-01-19

Component Code	Component Description	Component Lot Number
KS8000	KAPA Pure Beads (5 ml)	00066318

CoA's are not issued for complete kits, but for the individual component lots from which kits are assembled. CoA's for all component lots listed are attached.

Generated By	Date
Melissa Jonas (QC Scientist)	2016-10-19



Certificate of Analysis

PRODUCT DETAILS

Product name	KAPA Pure Beads	
Code & Lot number	KS8000	66318
Pack size	5 mL	
Bulk Code & Lot number	BS0002	66204

QUALITY CONTROL PARAMETERS

Parameter	Specification	Result
Functional assay	KAPA Pure Beads are functionally tested by capturing fragmented plasmid DNA at a 1.2X clean-up ratio. Fragment capture profile is determined by gel electrophoresis.	Passed
DNA contamination	A standard KAPA SYBR FAST no template reaction with KAPA Pure Beads (buffer only) contains <45 fg/μL bacterial genomic DNA (<i>E. coli</i> and related strains); as assessed by amplification of a 217 bp 16S rRNA fragment using a multicopy primer set in a 40-cycle reaction, no detectable human genomic DNA (as assessed by amplification of a 290 bp b-actin fragment using a multicopy primer set in a 40-cycle reaction) and no detectable adapter-ligated library DNA (as assessed by amplification using primers specific for the Illumina® TruSeq™ adapters or Ion Torrent™ adapters).	Passed

Generated by Toni Marinus (QC Scientist)

2016-04-07